





neodisher[®] Septo Fin



Disinfectant for the manual reprocessing of thermally stable and thermally instable instruments



Liquid concentrate

Fields of application:

- Manual disinfection of thermally stable and thermally instable instruments, including flexible endoscopes, TEE probes, MIS instruments and anaesthetic equipment

Performance spectrum:

- Bactericidal (incl. MRSA, H.pylori), mycobactericidal (incl. tuberculosis pathogen), fungicidal, virucidal and sporicidal (C. difficile) activity has been confirmed and certified
- Disinfecting activity has been tested and proved according to European standards, DGHM¹ and DVV/RKI² methods
- Low-foam formulation, suitable for use in circulation processes, e.g. semi-automatic machines
- Compatible with all neodisher process chemicals for reprocessing in washer disinfectors resp. AERs
- Suitable for materials such as stainless steel, chrome nickel steel, anodised aluminium, plastics (incl. silicone), rubber, porcelain and glass, non-ferrous metals
- VAH³-listed, ÖGHMP⁴-listed, on the list of the IHO⁵ for disinfectants with virucidal activity
- Pleasant odour, mildly perfumed

Special properties:

- Free of formaldehyde and glyoxal
- Excellent material compatibility
- Usable at all degrees of water hardness

Application and dosage:

neodisher Septo Fin can be used in immersion baths and semi-automatic machines. Depending on the desired activity a solution is prepared according to the application recommendations listed below.

Before disinfecting, the instruments resp. endoscopes are to be cleaned thoroughly, e.g. with the detergents neodisher MediClean forte or neodisher endo CLEAN. Then rinse the instruments resp. endoscopes thoroughly, immerse them in the neodisher Septo Fin-solution and adhere to the respective action time.

Instruments are to be dismantled and opened as far as possible. All surfaces must be completely wetted with the disinfectant solution. Air bubbles must be removed.

Disinfecting activity (20 °C, clean conditions)	
bactericidal, yeastocidal	30 ml/l (3.0 %), 15 min 20 ml/l (2.0 %), 30 min 10 ml/l (1.0 %), 60 min
mycobactericidal	50 ml/l (5.0 %), 30 min 40 ml/l (4.0 %), 60 min
fungicidal	50 ml/l (5.0 %), 60 min
C. difficile	70 ml/l (7.0 %), 6 hrs 40 ml/l (4.0 %), 24 hrs
active against enveloped viruses	10 ml/l (1.0 %), 15 min
virucidal	30 ml/l (3.0 %), 15 min 15 ml/l (1.5 %), 30 min

After expiry of the action time rinse the instruments thoroughly with running water of at least drinking water quality (preferably with deionised water). The stability of an unused working solution is 28 days. However, contaminated working solutions are to be renewed daily according to the RKI-guidelines.

Notes on application:

- For professional use only!
- It is generally recommended to wear gloves when working with disinfectants.
- For controlled dosing we recommend using manual dosing aids, such as e.g. dosing equipment for disinfectants. Please contact us.
- Do not mix with other products.
- Storage and transport is only permitted in original packaging.
- Reprocessing should comply with all ordinances pursuant to the Medical Device Directive and

With the above information, to our current knowledge we describe our product regarding safety necessities, but we do not involve any quality description or promise certain properties.



neodisher® Septo Fin

should be performed with appropriate validated processes.

- Please observe the instrument manufacturer's recommendations for use according to the requirements of DIN EN ISO 17664.
- Before using aldehyde-containing disinfectants such as neodisher Septo Fin no amine-containing products should be used as to avoid the formation of layers.

Expert reports:

The disinfecting activity has been confirmed by certification. We will be pleased to provide certificates on request.

Technical data:

pH-value	Approx. 4.0 (10 - 70 ml/l, determined in deionised water, 20 °C)
density	1.1g/cm ³ (20 °C)
viscosity	<50 mPa s (concentrate, 20 °C)


Ingredients:

Disinfecting substances in 100 g: 15 g glutaral

CE-mark: 0297

neodisher® Septo Fin conforms to the European directive 93/42/EEC, Annex I, concerning medical devices.

Storage information:

Always store at a temperature range between 0 °C and 30 °C. Usable for 3 years when stored as recommended. For expiry date refer to the stamp mark on the label behind the hour glass symbol .

Hazard and precautionary statements:

For safety information see EC safety data sheets. These are available e.g. at www.drweigert.de under the category "Service".

Dispose only when container is empty and closed. For disposal of product residues, refer to the Material Safety Data Sheet.

¹ Deutsche Gesellschaft für Mikrobiologie und Hygiene (German Society of Hygiene and Microbiology)

² Deutsche Vereinigung zur Bekämpfung der Viruskrankheiten (German Association for the Control of Virus Diseases) / Robert Koch Institute, Germany

³ Verbund für Angewandte Hygiene e.V. (Association of Applied Hygiene)

⁴ Österreichische Gesellschaft für Hygiene, Mikrobiologie und Präventivmedizin (Austrian Society of Hygiene, Microbiology and Preventive Medicine)

⁵ Industrieverband Hygiene und Oberflächenschutz (Industry Association Surface Protection and Hygiene, Germany)

MB 4207/3-1 date of issue 02/2014

With the above information, to our current knowledge we describe our product regarding safety necessities, but we do not involve any quality description or promise certain properties.